

计量授权证书附件

序号 No.	授权检定 项目名称 Item of Authorized Verification	测量范围 Measurement Range	准确度等级或 测量扩展不确定度 Accuracy or Expanded Measurement Uncertainty	依据检定 规程编号 Number of Verification Specification Referred to
1	pH 计检定仪	pH: 0~14 电位: (-2000~2000)mV	0.003 级、 0.0006 级	JJG 919
2	刮板细度计	(0~150) μm	MPE: $\pm (1.5 \sim 3.5) \mu\text{m}$	JJG 905
3	铁路支距尺	(100~1800) mm	MPE: $\pm 0.4 \text{ mm}$	JJG 1108
4	定碳定硫分析仪	C: (0.005~4.00) % S: (0.003~0.200) %	红外碳硫分析仪 C: (0.005~4.00) % MPE: $\pm(0.002 \sim 0.030)\%$ S: (0.003~0.200) % MPE: $\pm(0.001 \sim 0.010)\%$ 自动高速碳硫分析 仪 C: (0.030~4.00) % MPE: $\pm(0.010 \sim 0.050)\%$ S: (0.003~0.200) % MPE: $\pm(0.001 \sim 0.015)\%$	JJG 395

-----以下空白-----